

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100224\
 Data File : VW030332.D
 Acq On : 02 Oct 2024 17:17
 Operator : SY/MD
 Sample : P4237-05RE
 Misc : 4.94g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4F10RE

Quant Time: Oct 03 04:12:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 03 04:10:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	251301	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	141684	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	24928	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	59832	30.469	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 121.880%	
7) Chloroethane-d5	2.899	69	42805	29.959	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 119.840%	
11) 1,1-Dichloroethene-d2	4.021	65	34125	23.806	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 95.240%	
21) 2-Butanone-d5	7.081	46	9151	11.588	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 23.180%	
24) Chloroform-d	7.655	84	134350	19.900	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 79.600%	
26) 1,2-Dichloroethane-d4	8.307	65	71237	20.216	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 80.880%	
32) Benzene-d6	8.276	84	239368	30.143	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 120.560%	
36) 1,2-Dichloropropane-d6	9.276	67	77167	32.309	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 129.240%#	
41) Toluene-d8	10.325	98	169590	23.290	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 93.160%	
43) trans-1,3-Dichloroprop...	10.575	79	11031	11.104	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 44.400%	
47) 2-Hexanone-d5	10.922	63	5166	13.099	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 26.200%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	29672	16.626	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 66.520%	
66) 1,2-Dichlorobenzene-d4	13.849	152	14769	17.845	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 71.360%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100224\
Data File : VW030332.D
Acq On : 02 Oct 2024 17:17
Operator : SY/MD
Sample : P4237-05RE
Misc : 4.94g/10mL/MSVOA_W/SOIL/B
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4F10RE

Quant Time: Oct 03 04:12:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 03 04:10:12 2024
Response via : Initial Calibration

