

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100224\
 Data File : VW030350.D
 Acq On : 03 Oct 2024 02:04
 Operator : SY/MD
 Sample : P4238-11
 Misc : 5.48g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4EY7

Quant Time: Oct 03 05:12:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 03 05:10:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	343450	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	243377	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	61696	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	68352	25.468	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.880%
7) Chloroethane-d5	2.899	69	49909	25.559	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	102.240%
11) 1,1-Dichloroethene-d2	4.021	65	39046	19.931	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	79.720%
21) 2-Butanone-d5	7.081	46	24521	22.720	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	45.440%
24) Chloroform-d	7.654	84	160633	17.409	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	69.640%
26) 1,2-Dichloroethane-d4	8.307	65	80927	16.804	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	67.200%#
32) Benzene-d6	8.276	84	319688	23.437	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.760%
36) 1,2-Dichloropropane-d6	9.276	67	95071	23.173	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.680%
41) Toluene-d8	10.325	98	266108	21.275	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.120%
43) trans-1,3-Dichloroprop...	10.575	79	27778	16.278	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	65.120%
47) 2-Hexanone-d5	10.928	63	18290	26.998	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	54.000%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	46480	15.162	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	60.640%
66) 1,2-Dichlorobenzene-d4	13.848	152	37118	18.121	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	72.480%#
Target Compounds					Qvalue	
13) Acetone	4.131	43	6228	5.440	ug/L	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100224\
Data File : VW030350.D
Acq On : 03 Oct 2024 02:04
Operator : SY/MD
Sample : P4238-11
Misc : 5.48g/10mL/MSVOA_W/SOIL/A
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4EY7

Quant Time: Oct 03 05:12:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 03 05:10:23 2024
Response via : Initial Calibration

