

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100224\
 Data File : VW030355.D
 Acq On : 03 Oct 2024 04:04
 Operator : SY/MD
 Sample : P4238-16
 Misc : 5.46g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4F00

Quant Time: Oct 03 05:13:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 03 05:10:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	297516	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	242136	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	93189	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	61754	26.562	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	106.240%
7) Chloroethane-d5	2.899	69	46031	27.212	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	108.840%
11) 1,1-Dichloroethene-d2	4.021	65	36322	21.403	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	85.600%
21) 2-Butanone-d5	7.081	46	30303	32.412	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	64.820%
24) Chloroform-d	7.654	84	165323	20.684	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	82.720%
26) 1,2-Dichloroethane-d4	8.307	65	90692	21.739	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	86.960%
32) Benzene-d6	8.276	84	318475	23.467	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.880%
36) 1,2-Dichloropropane-d6	9.276	67	101564	24.883	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	99.520%
41) Toluene-d8	10.325	98	263952	21.211	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	84.840%
43) trans-1,3-Dichloroprop...	10.581	79	34016	20.036	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	80.160%
47) 2-Hexanone-d5	10.922	63	23112	34.290	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	68.580%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	64975	21.303	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	85.200%
66) 1,2-Dichlorobenzene-d4	13.848	152	64342	20.796	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	83.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100224\
Data File : VW030355.D
Acq On : 03 Oct 2024 04:04
Operator : SY/MD
Sample : P4238-16
Misc : 5.46g/10mL/MSVOA_W/SOIL/A
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4F00

Quant Time: Oct 03 05:13:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 03 05:10:23 2024
Response via : Initial Calibration

