

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100224\
 Data File : VW030357.D
 Acq On : 03 Oct 2024 04:52
 Operator : SY/MD
 Sample : P4238-18
 Misc : 4.97g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4F02

Quant Time: Oct 03 05:37:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 03 05:10:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.849	114	326749	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	245992	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	67401	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	71831	28.133	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 112.520%		
7) Chloroethane-d5	2.899	69	51406	27.671	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 110.680%		
11) 1,1-Dichloroethene-d2	4.027	65	40284	21.614	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	= 86.440%		
21) 2-Butanone-d5	7.081	46	23033	22.432	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 44.860%		
24) Chloroform-d	7.654	84	164405	18.729	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	= 74.920%		
26) 1,2-Dichloroethane-d4	8.307	65	83646	18.256	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	= 73.040%		
32) Benzene-d6	8.276	84	321870	23.346	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	= 93.400%		
36) 1,2-Dichloropropane-d6	9.276	67	97552	23.525	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	= 94.120%		
41) Toluene-d8	10.325	98	262074	20.730	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	= 82.920%		
43) trans-1,3-Dichloroprop...	10.581	79	28307	16.412	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	= 65.640%		
47) 2-Hexanone-d5	10.922	63	16435	24.002	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 48.000%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	50902	16.427	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	= 65.720%		
66) 1,2-Dichlorobenzene-d4	13.854	152	41421	18.510	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	= 74.040%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100224\
Data File : VW030357.D
Acq On : 03 Oct 2024 04:52
Operator : SY/MD
Sample : P4238-18
Misc : 4.97g/10mL/MSVOA_W/SOIL/A
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4F02

Quant Time: Oct 03 05:37:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 03 05:10:23 2024
Response via : Initial Calibration

