

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100224\
 Data File : VW030333.D
 Acq On : 02 Oct 2024 17:41
 Operator : SY/MD
 Sample : P4237-06
 Misc : 5.14g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4F11

Quant Time: Oct 03 04:13:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 03 04:10:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	298739	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	219530	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	49447	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	63571	27.232	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	108.920%	
7) Chloroethane-d5	2.899	69	44909	26.440	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	105.760%	
11) 1,1-Dichloroethene-d2	4.021	65	36363	21.339	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	85.360%	
21) 2-Butanone-d5	7.081	46	21480	22.881	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	45.760%	
24) Chloroform-d	7.655	84	147154	18.336	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	73.360%	
26) 1,2-Dichloroethane-d4	8.307	65	74632	17.816	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	71.280%	
32) Benzene-d6	8.276	84	283960	23.079	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	92.320%	
36) 1,2-Dichloropropane-d6	9.276	67	86600	23.401	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	93.600%	
41) Toluene-d8	10.325	98	233286	20.677	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	82.720%	
43) trans-1,3-Dichloroprop...	10.575	79	26886	17.467	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	69.880%	
47) 2-Hexanone-d5	10.922	63	16589	27.147	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	54.300%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	40619	14.689	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	58.760%	
66) 1,2-Dichlorobenzene-d4	13.849	152	29066	17.705	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	70.800%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100224\
Data File : VW030333.D
Acq On : 02 Oct 2024 17:41
Operator : SY/MD
Sample : P4237-06
Misc : 5.14g/10mL/MSVOA_W/SOIL/B
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4F11

Quant Time: Oct 03 04:13:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 03 04:10:12 2024
Response via : Initial Calibration

