

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100224\
 Data File : VW030342.D
 Acq On : 02 Oct 2024 21:16
 Operator : SY/MD
 Sample : P4238-04
 Misc : 4.85g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4F14

Quant Time: Oct 03 04:15:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 03 04:10:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	248200	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	154867	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	30026	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	53565	27.618	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	110.480%
7) Chloroethane-d5	2.899	69	38287	27.131	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	108.520%
11) 1,1-Dichloroethene-d2	4.021	65	30015	21.201	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	84.800%
21) 2-Butanone-d5	7.075	46	23155	29.687	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	59.380%
24) Chloroform-d	7.648	84	125018	18.749	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	75.000%
26) 1,2-Dichloroethane-d4	8.307	65	69643	20.010	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	80.040%
32) Benzene-d6	8.276	84	235522	27.134	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	108.520%
36) 1,2-Dichloropropane-d6	9.276	67	75381	28.875	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	115.480%
41) Toluene-d8	10.325	98	167986	21.106	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	84.440%
43) trans-1,3-Dichloroprop...	10.581	79	23459	21.604	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	86.400%
47) 2-Hexanone-d5	10.922	63	16688	38.711	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	77.420%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	40100	20.556	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	82.240%
66) 1,2-Dichlorobenzene-d4	13.848	152	20042	20.104	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	80.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100224\
Data File : VW030342.D
Acq On : 02 Oct 2024 21:16
Operator : SY/MD
Sample : P4238-04
Misc : 4.85g/10mL/MSVOA_W/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4F14

Quant Time: Oct 03 04:15:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 03 04:10:12 2024
Response via : Initial Calibration

