

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100224\
 Data File : VW030356.D
 Acq On : 03 Oct 2024 04:28
 Operator : SY/MD
 Sample : P4238-17
 Misc : 5.12g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4F01

Quant Time: Oct 03 05:14:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 03 05:10:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	269041	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	213417	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	80115	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	62278	29.623	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 118.480%		
7) Chloroethane-d5	2.899	69	44578	29.142	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 116.560%		
11) 1,1-Dichloroethene-d2	4.027	65	32689	21.301	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	= 85.200%		
21) 2-Butanone-d5	7.075	46	20902	24.723	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 49.440%		
24) Chloroform-d	7.654	84	143884	19.907	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	= 79.640%		
26) 1,2-Dichloroethane-d4	8.307	65	81355	21.565	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	= 86.240%		
32) Benzene-d6	8.276	84	266429	22.274	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	= 89.080%		
36) 1,2-Dichloropropane-d6	9.276	67	88590	24.625	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	= 98.480%		
41) Toluene-d8	10.325	98	198280	18.078	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	= 72.320%		
43) trans-1,3-Dichloroprop...	10.575	79	29476	19.698	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	= 78.800%		
47) 2-Hexanone-d5	10.922	63	16750	28.195	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 56.400%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	58998	21.946	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	= 87.800%		
66) 1,2-Dichlorobenzene-d4	13.848	152	52340	19.677	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	= 78.720%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100224\
Data File : VW030356.D
Acq On : 03 Oct 2024 04:28
Operator : SY/MD
Sample : P4238-17
Misc : 5.12g/10mL/MSVOA_W/SOIL/A
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4F01

Quant Time: Oct 03 05:14:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 03 05:10:23 2024
Response via : Initial Calibration

