

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100224\
 Data File : VW030358.D
 Acq On : 03 Oct 2024 05:15
 Operator : SY/MD
 Sample : P4238-19
 Misc : 5.07g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4F06

Quant Time: Oct 03 06:13:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 03 05:10:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	299000	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	245402	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	88261	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	62955	26.945	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 107.760%	
7) Chloroethane-d5	2.899	69	44424	26.132	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 104.520%	
11) 1,1-Dichloroethene-d2	4.021	65	34824	20.418	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 81.680%	
21) 2-Butanone-d5	7.075	46	23167	24.656	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 49.320%	
24) Chloroform-d	7.648	84	150992	18.797	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 75.200%	
26) 1,2-Dichloroethane-d4	8.307	65	82939	19.782	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 79.120%	
32) Benzene-d6	8.276	84	286384	20.822	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 83.280%	
36) 1,2-Dichloropropane-d6	9.276	67	92351	22.324	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 89.280%	
41) Toluene-d8	10.325	98	228113	18.087	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 72.360%	
43) trans-1,3-Dichloroprop...	10.575	79	23934	13.910	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 55.640%	
47) 2-Hexanone-d5	10.922	63	14798	21.663	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 43.320%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	58401	18.893	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 75.560%	
66) 1,2-Dichlorobenzene-d4	13.849	152	57069	19.475	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 77.920%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100224\
Data File : VW030358.D
Acq On : 03 Oct 2024 05:15
Operator : SY/MD
Sample : P4238-19
Misc : 5.07g/10mL/MSVOA_W/SOIL/A
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4F06

Quant Time: Oct 03 06:13:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 03 05:10:23 2024
Response via : Initial Calibration

