

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100225\
 Data File : VW032328.D
 Acq On : 02 Oct 2025 13:35
 Operator : SY/MD
 Sample : Q3257-02
 Misc : 11.59g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ENV2-6FT

Quant Time: Oct 02 21:37:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 02 21:24:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	197845	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	430890	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	442868	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	232780	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	205501	71.214	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	= 142.420%	
35) Dibromofluoromethane	7.904	113	159110	57.019	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	= 114.040%	
50) Toluene-d8	10.325	98	587929	57.086	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	= 114.180%	
62) 4-Bromofluorobenzene	12.617	95	226278	58.129	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	= 116.260%	
Target Compounds						Qvalue
16) Acetone	4.155	43	27330	34.204	ug/l	99
20) Methylene Chloride	4.960	84	26183	5.098	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100225\
Data File : VW032328.D
Acq On : 02 Oct 2025 13:35
Operator : SY/MD
Sample : Q3257-02
Misc : 11.59g/5mL/MSVOA_W/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
ENV2-6FT

Quant Time: Oct 02 21:37:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 02 21:24:16 2025
Response via : Initial Calibration

