

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100225\
 Data File : VW032329.D
 Acq On : 02 Oct 2025 13:57
 Operator : SY/MD
 Sample : Q3266-01
 Misc : 14.74g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 ENV-3-6FT

Quant Time: Oct 02 21:37:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 02 21:24:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	180169	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	409718	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	401656	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	185918	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	194570	74.041	ug/l	0.00
Spiked Amount 50.000	Range 63 - 155		Recovery = 148.080%			
35) Dibromofluoromethane	7.904	113	149792	56.453	ug/l	0.00
Spiked Amount 50.000	Range 70 - 134		Recovery = 112.900%			
50) Toluene-d8	10.324	98	535418	54.673	ug/l	0.00
Spiked Amount 50.000	Range 74 - 123		Recovery = 109.340%			
62) 4-Bromofluorobenzene	12.617	95	200173	54.080	ug/l	0.00
Spiked Amount 50.000	Range 17 - 146		Recovery = 108.160%			
Target Compounds					Qvalue	
16) Acetone	4.161	43	68048	93.519	ug/l	96
20) Methylene Chloride	4.966	84	19304	2.987	ug/l	94
25) 2-Butanone	7.197	43	19208	19.382	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100225\
Data File : VW032329.D
Acq On : 02 Oct 2025 13:57
Operator : SY/MD
Sample : Q3266-01
Misc : 14.74g/5mL/MSVOA_W/SOIL/A
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
ENV-3-6FT

Quant Time: Oct 02 21:37:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 02 21:24:16 2025
Response via : Initial Calibration

