

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100225\
 Data File : VW032336.D
 Acq On : 02 Oct 2025 16:31
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK

Quant Time: Oct 02 22:12:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 02 21:24:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	212138	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	456706	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	463790	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	243745	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	174834	56.504	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	113.000%
35) Dibromofluoromethane	7.904	113	149713	50.618	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	101.240%
50) Toluene-d8	10.324	98	579239	53.063	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	106.120%
62) 4-Bromofluorobenzene	12.617	95	223878	54.261	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	108.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100225\
Data File : VW032336.D
Acq On : 02 Oct 2025 16:31
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK

Quant Time: Oct 02 22:12:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 02 21:24:16 2025
Response via : Initial Calibration

