

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100324\
 Data File : VW030369.D
 Acq On : 03 Oct 2024 14:21
 Operator : SY/MD
 Sample : P4238-05
 Misc : 4.75g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4F15

Quant Time: Oct 04 01:08:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 04 01:05:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	264230	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	188607	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	55164	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	61376	29.725	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 118.920%		
7) Chloroethane-d5	2.899	69	43739	29.114	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 116.440%		
11) 1,1-Dichloroethene-d2	4.021	65	32724	21.712	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	= 86.840%		
21) 2-Butanone-d5	7.075	46	23262	28.015	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 56.040%		
24) Chloroform-d	7.648	84	143343	20.193	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	= 80.760%		
26) 1,2-Dichloroethane-d4	8.307	65	76038	20.522	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	= 82.080%		
32) Benzene-d6	8.276	84	263173	24.896	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	= 99.600%		
36) 1,2-Dichloropropane-d6	9.270	67	83983	26.415	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	= 105.680%		
41) Toluene-d8	10.325	98	201050	20.741	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	= 82.960%		
43) trans-1,3-Dichloroprop...	10.575	79	24411	18.459	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	= 73.840%		
47) 2-Hexanone-d5	10.922	63	16528	31.482	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 62.960%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	44946	18.919	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	= 75.680%		
66) 1,2-Dichlorobenzene-d4	13.848	152	29440	16.074	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	= 64.280%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100324\
Data File : VW030369.D
Acq On : 03 Oct 2024 14:21
Operator : SY/MD
Sample : P4238-05
Misc : 4.75g/10mL/MSVOA_W/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4F15

Quant Time: Oct 04 01:08:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 04 01:05:34 2024
Response via : Initial Calibration

