

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100324\
 Data File : VW030365.D
 Acq On : 03 Oct 2024 11:56
 Operator : SY/MD
 Sample : P4238-01RE
 Misc : 5.34g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4EQ8RE

Quant Time: Oct 04 01:07:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 04 01:05:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	391370	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	355246	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	145781	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	66276	21.671	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	86.680%	
7) Chloroethane-d5	2.905	69	48679	21.876	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	87.520%	
11) 1,1-Dichloroethene-d2	4.021	65	37272	16.696	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	66.800%	
21) 2-Butanone-d5	7.075	46	33142	26.947	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	53.900%	
24) Chloroform-d	7.648	84	170212	16.189	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	64.760%	
26) 1,2-Dichloroethane-d4	8.307	65	89526	16.313	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	65.240%#	
32) Benzene-d6	8.276	84	322991	16.222	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	64.880%	
36) 1,2-Dichloropropane-d6	9.276	67	102928	17.188	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	68.760%#	
41) Toluene-d8	10.325	98	291474	15.965	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	63.840%	
43) trans-1,3-Dichloroprop...	10.581	79	39003	15.659	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	62.640%	
47) 2-Hexanone-d5	10.922	63	25957	26.249	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	52.500%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	66973	14.967	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	59.880%	
66) 1,2-Dichlorobenzene-d4	13.848	152	86260	17.822	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	71.280%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100324\
Data File : VW030365.D
Acq On : 03 Oct 2024 11:56
Operator : SY/MD
Sample : P4238-01RE
Misc : 5.34g/10mL/MSVOA_W/SOIL/B
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4EQ8RE

Quant Time: Oct 04 01:07:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 04 01:05:34 2024
Response via : Initial Calibration

