

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100422\
 Data File : VW024723.D
 Acq On : 04 Oct 2022 18:44
 Operator : SY/VA
 Sample : N4973-10
 Misc : 6.57g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EY1A7

Quant Time: Oct 05 07:19:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Sep 30 22:12:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	584977	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	448085	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	141759	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	87991	14.377	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	57.520%
7) Chloroethane-d5	2.885	69	54349	15.281	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.120%
11) 1,1-Dichloroethene-d2	4.019	63	116399	8.217	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	32.880%#
21) 2-Butanone-d5	7.080	46	142082	64.570	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	129.140%
24) Chloroform-d	7.647	84	335602	21.460	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.840%
26) 1,2-Dichloroethane-d4	8.305	65	178281	21.976	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.920%
32) Benzene-d6	8.275	84	546690	19.088	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	76.360%
36) 1,2-Dichloropropane-d6	9.274	67	209808	23.976	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	95.920%
41) Toluene-d8	10.323	98	382376	14.938	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	59.760%
43) trans-1,3-Dichloroprop...	10.579	79	58670	15.056	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	60.240%
47) 2-Hexanone-d5	10.920	63	100725	67.276	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	134.560%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	199206	31.950	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	127.800%#
66) 1,2-Dichlorobenzene-d4	13.853	152	113633	22.780	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.120%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100422\
Data File : VW024723.D
Acq On : 04 Oct 2022 18:44
Operator : SY/VA
Sample : N4973-10
Misc : 6.57g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EY1A7

Quant Time: Oct 05 07:19:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Sep 30 22:12:13 2022
Response via : Initial Calibration

