

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100424\
 Data File : VW030387.D
 Acq On : 04 Oct 2024 12:41
 Operator : SY/MD
 Sample : VW1004SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Oct 05 01:26:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Oct 05 01:25:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	188916	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	183068	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	80045	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	39161	26.528	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	106.120%
7) Chloroethane-d5	2.899	69	31031	28.890	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	115.560%
11) 1,1-Dichloroethene-d2	4.021	65	24625	22.852	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	91.400%
21) 2-Butanone-d5	7.081	46	26739	45.041	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.080%
24) Chloroform-d	7.648	84	120149	23.674	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.680%
26) 1,2-Dichloroethane-d4	8.307	65	64117	24.204	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	96.800%
32) Benzene-d6	8.276	84	232664	22.676	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	90.720%
36) 1,2-Dichloropropane-d6	9.276	67	74044	23.994	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	95.960%
41) Toluene-d8	10.325	98	206858	21.986	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.960%
43) trans-1,3-Dichloroprop...	10.581	79	25766	20.073	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	80.280%
47) 2-Hexanone-d5	10.922	63	21330	41.857	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	83.720%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	52773	22.885	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	91.560%
66) 1,2-Dichlorobenzene-d4	13.848	152	67152	25.268	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	101.080%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100424\
Data File : VW030387.D
Acq On : 04 Oct 2024 12:41
Operator : SY/MD
Sample : VW1004SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Oct 05 01:26:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Oct 05 01:25:07 2024
Response via : Initial Calibration

