

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100424\
 Data File : VW030397.D
 Acq On : 04 Oct 2024 16:47
 Operator : SY/MD
 Sample : P4239-19
 Misc : 4.85g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4EP8

Quant Time: Oct 05 01:28:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Oct 05 01:25:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	376396	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	324230	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	122782	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	61094	20.771	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.080%
7) Chloroethane-d5	2.899	69	47176	22.044	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.160%
11) 1,1-Dichloroethene-d2	4.021	65	37126	17.292	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.160%
21) 2-Butanone-d5	7.081	46	34279	28.981	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	57.960%
24) Chloroform-d	7.648	84	185933	18.388	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	73.560%
26) 1,2-Dichloroethane-d4	8.307	65	92528	17.531	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	70.120%
32) Benzene-d6	8.276	84	350209	19.272	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	77.080%
36) 1,2-Dichloropropane-d6	9.270	67	110448	20.208	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	80.840%
41) Toluene-d8	10.325	98	302992	18.183	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	72.720%
43) trans-1,3-Dichloroprop...	10.581	79	39826	17.519	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	70.080%
47) 2-Hexanone-d5	10.922	63	27954	30.973	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	61.940%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	71951	17.617	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	70.480%
66) 1,2-Dichlorobenzene-d4	13.848	152	78181	19.178	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	76.720%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100424\
Data File : VW030397.D
Acq On : 04 Oct 2024 16:47
Operator : SY/MD
Sample : P4239-19
Misc : 4.85g/10mL/MSVOA_W/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4EP8

Quant Time: Oct 05 01:28:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Oct 05 01:25:07 2024
Response via : Initial Calibration

