

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100622\
 Data File : VW024761.D
 Acq On : 06 Oct 2022 15:54
 Operator : SY/VA
 Sample : N5023-02
 Misc : 6.56g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0BD5

Quant Time: Oct 07 03:07:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 07 03:04:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	489269	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	403057	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	150843	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	74626	14.578	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	58.320%		
7) Chloroethane-d5	2.885	69	50696	17.042	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	68.160%		
11) 1,1-Dichloroethene-d2	4.019	63	105281	8.886	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery =	35.560%#		
21) 2-Butanone-d5	7.086	46	109966	59.750	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	119.500%		
24) Chloroform-d	7.647	84	274640	20.997	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery =	84.000%		
26) 1,2-Dichloroethane-d4	8.305	65	155145	22.865	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery =	91.440%		
32) Benzene-d6	8.275	84	436854	16.957	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery =	67.840%		
36) 1,2-Dichloropropane-d6	9.274	67	175738	22.327	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery =	89.320%		
41) Toluene-d8	10.323	98	314134	13.643	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery =	54.560%		
43) trans-1,3-Dichloroprop...	10.579	79	54019	15.411	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery =	61.640%		
47) 2-Hexanone-d5	10.927	63	83169	61.756	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	123.520%		
56) 1,1,2,2-Tetrachloroeth...	12.689	84	174089	31.040	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery =	124.160%#		
66) 1,2-Dichlorobenzene-d4	13.847	152	101554	19.132	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery =	76.520%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100622\
 Data File : VW024761.D
 Acq On : 06 Oct 2022 15:54
 Operator : SY/VA
 Sample : N5023-02
 Misc : 6.56g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0BD5

Quant Time: Oct 07 03:07:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 07 03:04:02 2022
 Response via : Initial Calibration

