

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100622\
 Data File : VW024754.D
 Acq On : 06 Oct 2022 13:06
 Operator : SY/VA
 Sample : N4991-02RE
 Misc : 4.64g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0BE9RE

Quant Time: Oct 07 03:06:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 07 03:04:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	502254	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	379446	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	104751	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	95073	18.092	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.360%
7) Chloroethane-d5	2.891	69	64085	20.986	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.960%
11) 1,1-Dichloroethene-d2	4.019	63	139163	11.442	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	45.760%
21) 2-Butanone-d5	7.086	46	158059	83.661	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	167.320%#
24) Chloroform-d	7.653	84	384379	28.627	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	114.520%
26) 1,2-Dichloroethane-d4	8.311	65	207201	29.747	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	119.000%
32) Benzene-d6	8.275	84	534677	22.046	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	88.200%
36) 1,2-Dichloropropane-d6	9.274	67	238064	32.127	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	128.520%#
41) Toluene-d8	10.323	98	369335	17.039	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	68.160%
43) trans-1,3-Dichloroprop...	10.579	79	70446	21.348	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.400%
47) 2-Hexanone-d5	10.927	63	110095	86.836	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	173.680%#
56) 1,1,2,2-Tetrachloroeth...	12.688	84	232410	44.018	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	176.080%#
66) 1,2-Dichlorobenzene-d4	13.853	152	84297	22.869	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.480%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100622\
Data File : VW024754.D
Acq On : 06 Oct 2022 13:06
Operator : SY/VA
Sample : N4991-02RE
Misc : 4.64g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0BE9RE

Quant Time: Oct 07 03:06:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 07 03:04:02 2022
Response via : Initial Calibration

