

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100625\
 Data File : VW032353.D
 Acq On : 06 Oct 2025 12:37
 Operator : SY/MD
 Sample : VW1006SBL01
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW1006SBL01

Quant Time: Oct 07 03:42:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 02 21:24:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	168139	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	378247	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	382025	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	188486	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	158534	64.644	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	129.280%
35) Dibromofluoromethane	7.898	113	137927	56.307	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	112.620%
50) Toluene-d8	10.325	98	472361	52.248	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	104.500%
62) 4-Bromofluorobenzene	12.617	95	169464	49.592	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	99.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100625\
Data File : VW032353.D
Acq On : 06 Oct 2025 12:37
Operator : SY/MD
Sample : VW1006SBL01
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW1006SBL01

Quant Time: Oct 07 03:42:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 02 21:24:16 2025
Response via : Initial Calibration

