

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100625\
 Data File : VW032356.D
 Acq On : 06 Oct 2025 14:05
 Operator : SY/MD
 Sample : Q3270-01
 Misc : 5.36g/5mL/MSVOA_W/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SG-1-TEST-PIT

Quant Time: Oct 07 03:45:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 02 21:24:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	8436	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	15244	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.635	117	13569	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	5535	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	7841	63.725	ug/l	0.00
Spiked Amount	50.000	Range 63 - 155	Recovery	=	127.460%	
35) Dibromofluoromethane	7.898	113	6115	61.942	ug/l	0.00
Spiked Amount	50.000	Range 70 - 134	Recovery	=	123.880%	
50) Toluene-d8	10.331	98	16837	46.210	ug/l	0.00
Spiked Amount	50.000	Range 74 - 123	Recovery	=	92.420%	
62) 4-Bromofluorobenzene	12.623	95	5753	41.774	ug/l	0.00
Spiked Amount	50.000	Range 17 - 146	Recovery	=	83.540%	
Target Compounds						
16) Acetone	4.179	43	2470m	72.498	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100625\
Data File : VW032356.D
Acq On : 06 Oct 2025 14:05
Operator : SY/MD
Sample : Q3270-01
Misc : 5.36g/5mL/MSVOA_W/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SG-1-TEST-PIT

Quant Time: Oct 07 03:45:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 02 21:24:16 2025
Response via : Initial Calibration

