

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100625\
 Data File : VW032360.D
 Acq On : 06 Oct 2025 15:32
 Operator : SY/MD
 Sample : Q3289-03
 Misc : 4.68g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VNJ-261

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 10/07/2025
 Supervised By :Semsettin Yesilyurt 10/07/2025

Quant Time: Oct 07 03:47:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 02 21:24:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	178194	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	392275	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	404257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	213244	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	170519	65.608	ug/l	0.00
Spiked Amount 50.000	Range 63 - 155		Recovery = 131.220%			
35) Dibromofluoromethane	7.904	113	143535	56.501	ug/l	0.00
Spiked Amount 50.000	Range 70 - 134		Recovery = 113.000%			
50) Toluene-d8	10.331	98	505222	53.884	ug/l	0.00
Spiked Amount 50.000	Range 74 - 123		Recovery = 107.760%			
62) 4-Bromofluorobenzene	12.617	95	184257	51.993	ug/l	0.00
Spiked Amount 50.000	Range 17 - 146		Recovery = 103.980%			
Target Compounds						
16) Acetone	4.167	43	10612	14.746	ug/l	97
18) Methyl Acetate	4.722	43	4542m	2.203	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100625\
Data File : VW032360.D
Acq On : 06 Oct 2025 15:32
Operator : SY/MD
Sample : Q3289-03
Misc : 4.68g/5mL/MSVOA_W/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VNJ-261

Quant Time: Oct 07 03:47:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 02 21:24:16 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 10/07/2025
Supervised By :Semsettin Yesilyurt 10/07/2025

