

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100625\
 Data File : VW032363.D
 Acq On : 06 Oct 2025 16:39
 Operator : SY/MD
 Sample : Q3295-03
 Misc : 5.28g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EO-02-10062025

Quant Time: Oct 07 03:49:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 02 21:24:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	191766	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	430557	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.636	117	449428	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	225653	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	174735	62.472	ug/l	0.00
Spiked Amount	50.000	Range 63 - 155	Recovery	= 124.940%		
35) Dibromofluoromethane	7.904	113	149056	53.457	ug/l	0.00
Spiked Amount	50.000	Range 70 - 134	Recovery	= 106.920%		
50) Toluene-d8	10.325	98	549452	53.391	ug/l	0.00
Spiked Amount	50.000	Range 74 - 123	Recovery	= 106.780%		
62) 4-Bromofluorobenzene	12.617	95	223954	57.576	ug/l	0.00
Spiked Amount	50.000	Range 17 - 146	Recovery	= 115.160%		
Target Compounds						
18) Methyl Acetate	4.716	43	4226	1.905	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100625\
Data File : VW032363.D
Acq On : 06 Oct 2025 16:39
Operator : SY/MD
Sample : Q3295-03
Misc : 5.28g/5mL/MSVOA_W/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EO-02-10062025

Quant Time: Oct 07 03:49:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 02 21:24:16 2025
Response via : Initial Calibration

