



Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100625\
Data File : VW032351.D
Acq On : 06 Oct 2025 11:18
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
Title : SW846 8260
Last Update : Thu Oct 02 21:24:16 2025



AutoFind: Scans 1759, 1760, 1761; Background Corrected with Scan 1751

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.9	13636	PASS
75	95	30	60	52.5	36043	PASS
95	95	100	100	100.0	68629	PASS
96	95	5	9	6.3	4300	PASS
173	174	0.00	2	1.0	466	PASS
174	95	50	100	70.0	48032	PASS
175	174	5	9	7.4	3533	PASS
176	174	95	101	97.6	46877	PASS
177	176	5	9	6.0	2799	PASS