

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100722\
 Data File : VW024775.D
 Acq On : 07 Oct 2022 10:51
 Operator : SY/VA
 Sample : N4973-15
 Misc : 4.67g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EY1A2

Quant Time: Oct 07 23:18:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 07 03:04:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	570342	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	518077	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	257044	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	84516	14.163	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	56.640%
7) Chloroethane-d5	2.891	69	59474	17.151	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	68.600%
11) 1,1-Dichloroethene-d2	4.019	63	132525	9.595	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	38.400%#
21) 2-Butanone-d5	7.080	46	123047	57.354	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	114.700%
24) Chloroform-d	7.653	84	334981	21.970	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.880%
26) 1,2-Dichloroethane-d4	8.305	65	173003	21.873	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.480%
32) Benzene-d6	8.274	84	542800	16.392	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	65.560%
36) 1,2-Dichloropropane-d6	9.274	67	207596	20.519	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	82.080%
41) Toluene-d8	10.323	98	410281	13.863	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	55.440%
43) trans-1,3-Dichloroprop...	10.579	79	68913	15.295	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	61.160%
47) 2-Hexanone-d5	10.926	63	91470	52.841	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	105.680%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	197545	27.403	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	109.600%
66) 1,2-Dichlorobenzene-d4	13.853	152	180121	19.914	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	79.640%
Target Compounds						
					Qvalue	
46) Tetrachloroethene	10.859	164	43392	6.949	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100722\
Data File : VW024775.D
Acq On : 07 Oct 2022 10:51
Operator : SY/VA
Sample : N4973-15
Misc : 4.67g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EY1A2

Quant Time: Oct 07 23:18:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 07 03:04:02 2022
Response via : Initial Calibration

