

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100722\
 Data File : VW024792.D
 Acq On : 07 Oct 2022 17:40
 Operator : SY/VA
 Sample : N5049-01
 Misc : 6.80g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0BG3

Quant Time: Oct 07 23:20:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 07 03:04:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	485209	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	426863	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	168831	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	60803	11.977	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	47.920%
7) Chloroethane-d5	2.885	69	45176	15.314	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.240%
11) 1,1-Dichloroethene-d2	4.019	63	90245	7.681	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	30.720%#
21) 2-Butanone-d5	7.086	46	119679	65.572	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	131.140%
24) Chloroform-d	7.653	84	250238	19.291	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	77.160%
26) 1,2-Dichloroethane-d4	8.305	65	145044	21.555	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	86.240%
32) Benzene-d6	8.275	84	389172	14.264	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	57.040%
36) 1,2-Dichloropropane-d6	9.274	67	158053	18.960	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	75.840%
41) Toluene-d8	10.323	98	287974	11.809	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	47.240%
43) trans-1,3-Dichloroprop...	10.579	79	46657	12.568	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	50.280%
47) 2-Hexanone-d5	10.926	63	83311	58.411	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	116.820%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	180536	30.395	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	121.560%#
66) 1,2-Dichlorobenzene-d4	13.847	152	105301	17.725	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	70.880%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100722\
Data File : VW024792.D
Acq On : 07 Oct 2022 17:40
Operator : SY/VA
Sample : N5049-01
Misc : 6.80g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0BG3

Quant Time: Oct 07 23:20:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 07 03:04:02 2022
Response via : Initial Calibration

