

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100725\
 Data File : VW032371.D
 Acq On : 07 Oct 2025 10:08
 Operator : SY/MD
 Sample : VW1007SBL01
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW1007SBL01

Quant Time: Oct 08 03:18:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 02 21:24:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	160398	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.849	114	388212	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	378174	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	179103	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.313	65	157305	67.239	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	134.480%
35) Dibromofluoromethane	7.898	113	138894	55.246	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	110.500%
50) Toluene-d8	10.325	98	475159	51.208	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	102.420%
62) 4-Bromofluorobenzene	12.617	95	168271	47.979	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	95.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100725\
Data File : VW032371.D
Acq On : 07 Oct 2025 10:08
Operator : SY/MD
Sample : VW1007SBL01
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW1007SBL01

Quant Time: Oct 08 03:18:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 02 21:24:16 2025
Response via : Initial Calibration

