

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100725\
 Data File : VW032370.D
 Acq On : 07 Oct 2025 09:25
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 08 03:16:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 02 21:24:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.314	0.337	-7.3	116	0.00
3 P	Chloromethane	0.484	0.458	5.4	100	0.00
4 C	Vinyl Chloride	0.578	0.583	-0.9#	100	0.00
5 T	Bromomethane	0.425	0.422	0.7	101	0.00
6 T	Chloroethane	0.382	0.398	-4.2	102	0.00
7 T	Trichlorofluoromethane	0.433	0.474	-9.5	103	0.00
8 T	Diethyl Ether	0.375	0.382	-1.9	101	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.519	0.537	-3.5	105	0.00
10 T	Methyl Iodide	0.741	0.769	-3.8	101	0.00
11 T	Tert butyl alcohol	0.055	0.055	0.0	102	0.00
12 CM	1,1-Dichloroethene	0.564	0.580	-2.8#	101	0.00
13 T	Acrolein	0.090	0.057	36.7#	65	0.00
14 T	Allyl chloride	0.952	0.938	1.5	94	0.00
15 T	Acrylonitrile	0.195	0.207	-6.2	101	0.00
16 T	Acetone	0.202	0.196	3.0	98	0.00
17 T	Carbon Disulfide	1.538	1.577	-2.5	100	0.00
18 T	Methyl Acetate	0.578	0.592	-2.4	86	0.00
19 T	Methyl tert-butyl Ether	1.110	1.122	-1.1	96	0.00
20 T	Methylene Chloride	0.838	0.673	19.7	97	0.00
21 T	trans-1,2-Dichloroethene	0.624	0.660	-5.8	102	0.00
22 T	Diisopropyl ether	1.979	2.075	-4.9	100	0.00
23 T	Vinyl Acetate	1.322	1.416	-7.1	103	0.00
24 P	1,1-Dichloroethane	1.203	1.254	-4.2	101	0.00
25 T	2-Butanone	0.275	0.275	0.0	96	0.00
26 T	2,2-Dichloropropane	0.615	0.641	-4.2	101	0.00
27 T	cis-1,2-Dichloroethene	0.749	0.778	-3.9	99	0.00
28 T	Bromochloromethane	0.566	0.581	-2.7	97	0.00
29 T	Tetrahydrofuran	0.173	0.186	-7.5	99	0.00
30 C	Chloroform	1.220	1.305	-7.0#	104	0.00
31 T	Cyclohexane	1.027	1.024	0.3	102	0.00
32 T	1,1,1-Trichloroethane	0.864	0.898	-3.9	100	0.00
33 S	1,2-Dichloroethane-d4	0.729	0.757	-3.8	99	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.324	0.339	-4.6	99	0.00
36 T	1,1-Dichloropropene	0.457	0.502	-9.8	102	0.00
37 T	Ethyl Acetate	0.304	0.335	-10.2	103	0.00
38 T	Carbon Tetrachloride	0.430	0.467	-8.6	104	0.00
39 T	Methylcyclohexane	0.551	0.596	-8.2	102	0.00
40 TM	Benzene	1.432	1.535	-7.2	102	0.00
41 T	Methacrylonitrile	0.164	0.175	-6.7	96	0.00
42 TM	1,2-Dichloroethane	0.469	0.506	-7.9	102	0.00
43 T	Isopropyl Acetate	0.540	0.584	-8.1	100	0.00
44 TM	Trichloroethene	0.335	0.351	-4.8	101	0.00
45 C	1,2-Dichloropropane	0.358	0.389	-8.7#	102	0.00
46 T	Dibromomethane	0.220	0.236	-7.3	100	0.00
47 T	Bromodichloromethane	0.503	0.551	-9.5	103	0.00
48 T	Methyl methacrylate	0.262	0.288	-9.9	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100725\
 Data File : VW032370.D
 Acq On : 07 Oct 2025 09:25
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 08 03:16:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 02 21:24:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.004	-33.3#	103	0.00
50 S	Toluene-d8	1.195	1.288	-7.8	97	0.00
51 T	4-Methyl-2-Pentanone	0.313	0.353	-12.8	101	0.00
52 CM	Toluene	0.892	0.970	-8.7#	101	0.00
53 T	t-1,3-Dichloropropene	0.488	0.520	-6.6	97	0.00
54 T	cis-1,3-Dichloropropene	0.562	0.609	-8.4	100	0.00
55 T	1,1,2-Trichloroethane	0.294	0.327	-11.2	104	0.00
56 T	Ethyl methacrylate	0.411	0.459	-11.7	101	0.00
57 T	1,3-Dichloropropane	0.519	0.576	-11.0	103	0.00
58 T	2-Chloroethyl Vinyl ether	0.219	0.259	-18.3	96	0.00
59 T	2-Hexanone	0.220	0.246	-11.8	100	0.00
60 T	Dibromochloromethane	0.334	0.378	-13.2	104	0.00
61 T	1,2-Dibromoethane	0.287	0.309	-7.7	100	0.00
62 S	4-Bromofluorobenzene	0.452	0.472	-4.4	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64 T	Tetrachloroethene	0.299	0.308	-3.0	99	0.00
65 PM	Chlorobenzene	1.100	1.166	-6.0	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.343	0.384	-12.0	108	0.00
67 C	Ethyl Benzene	1.834	1.985	-8.2#	102	0.00
68 T	m/p-Xylenes	0.707	0.785	-11.0	105	0.00
69 T	o-Xylene	0.661	0.710	-7.4	100	0.00
70 T	Styrene	1.155	1.302	-12.7	104	0.00
71 P	Bromoform	0.196	0.212	-8.2	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	3.421	3.794	-10.9	103	0.00
74 T	N-amyl acetate	1.091	1.223	-12.1	104	0.00
75 P	1,1,2,2-Tetrachloroethane	0.849	0.939	-10.6	105	0.00
76 T	1,2,3-Trichloropropane	0.658	0.694	-5.5	103	0.00
77 T	Bromobenzene	0.825	0.930	-12.7	103	0.00
78 T	n-propylbenzene	4.273	4.873	-14.0	107	0.00
79 T	2-Chlorotoluene	2.613	2.870	-9.8	103	0.00
80 T	1,3,5-Trimethylbenzene	2.958	3.300	-11.6	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.260	0.309	-18.8	104	0.00
82 T	4-Chlorotoluene	2.778	3.068	-10.4	104	0.00
83 T	tert-Butylbenzene	2.427	2.688	-10.8	103	0.00
84 T	1,2,4-Trimethylbenzene	3.008	3.431	-14.1	104	0.00
85 T	sec-Butylbenzene	3.709	4.045	-9.1	102	0.00
86 T	p-Isopropyltoluene	3.077	3.504	-13.9	105	0.00
87 T	1,3-Dichlorobenzene	1.670	1.846	-10.5	104	0.00
88 T	1,4-Dichlorobenzene	1.670	1.810	-8.4	105	0.00
89 T	n-Butylbenzene	3.005	3.388	-12.7	106	0.00
90 T	Hexachloroethane	0.552	0.623	-12.9	108	0.00
91 T	1,2-Dichlorobenzene	1.509	1.666	-10.4	107	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.149	0.160	-7.4	99	0.00
93 T	1,2,4-Trichlorobenzene	0.905	0.928	-2.5	100	0.00
94 T	Hexachlorobutadiene	0.414	0.469	-13.3	117	0.00
95 T	Naphthalene	2.174	2.445	-12.5	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100725\
Data File : VW032370.D
Acq On : 07 Oct 2025 09:25
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Oct 08 03:16:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 02 21:24:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.853	0.871	-2.1	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6