



Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100725\
Data File : VW032369.D
Acq On : 07 Oct 2025 08:42
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W100125S.M
Title : SW846 8260
Last Update : Thu Oct 02 21:24:16 2025



AutoFind: Scans 1759, 1760, 1761; Background Corrected with Scan 1751

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.9	20806	PASS
75	95	30	60	52.2	54715	PASS
95	95	100	100	100.0	104803	PASS
96	95	5	9	6.6	6896	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	68.8	72144	PASS
175	174	5	9	8.1	5852	PASS
176	174	95	101	95.7	69027	PASS
177	176	5	9	6.4	4418	PASS