

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100824\
 Data File : VW030462.D
 Acq On : 08 Oct 2024 21:23
 Operator : SY/MD
 Sample : P4329-21
 Misc : 5.59g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4FB8

Quant Time: Oct 09 01:22:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Oct 09 01:13:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	222057	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	186327	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	84606	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	27711	15.970	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	63.880%
7) Chloroethane-d5	2.905	69	22134	17.531	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	70.120%
11) 1,1-Dichloroethene-d2	4.021	65	22033	17.395	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.560%
21) 2-Butanone-d5	7.081	46	10688	15.316	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	30.640%
24) Chloroform-d	7.655	84	102840	17.239	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	68.960%
26) 1,2-Dichloroethane-d4	8.307	65	50764	16.303	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	65.200%#
32) Benzene-d6	8.276	84	206644	19.788	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	79.160%
36) 1,2-Dichloropropane-d6	9.276	67	53433	17.012	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	68.040%#
41) Toluene-d8	10.325	98	200928	20.982	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	83.920%
43) trans-1,3-Dichloroprop...	10.581	79	20613	15.778	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	63.120%
47) 2-Hexanone-d5	10.922	63	11230	21.652	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	43.300%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	34708	14.788	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	59.160%
66) 1,2-Dichlorobenzene-d4	13.849	152	58469	20.815	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	83.240%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100824\
Data File : VW030462.D
Acq On : 08 Oct 2024 21:23
Operator : SY/MD
Sample : P4329-21
Misc : 5.59g/10mL/MSVOA_W/SOIL/A
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4FB8

Quant Time: Oct 09 01:22:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Oct 09 01:13:46 2024
Response via : Initial Calibration

