

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100824\
 Data File : VW030465.D
 Acq On : 08 Oct 2024 23:23
 Operator : SY/MD
 Sample : VW1008SBL02
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK567

Quant Time: Oct 09 04:19:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Oct 09 01:13:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	120164	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	110709	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	55379	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	11508	12.256	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	49.040%	
7) Chloroethane-d5	2.899	69	10198	14.927	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	59.720%	
11) 1,1-Dichloroethene-d2	4.021	65	9983	14.565	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	58.240%	
21) 2-Butanone-d5	7.088	46	10021	26.538	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	53.080%	
24) Chloroform-d	7.648	84	56327	17.448	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	69.800%	
26) 1,2-Dichloroethane-d4	8.307	65	30158	17.898	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	71.600%	
32) Benzene-d6	8.276	84	107659	17.351	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	69.400%	
36) 1,2-Dichloropropane-d6	9.276	67	29239	15.667	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	62.680%#	
41) Toluene-d8	10.325	98	105081	18.469	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	73.880%	
43) trans-1,3-Dichloroprop...	10.575	79	12536	16.150	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	64.600%	
47) 2-Hexanone-d5	10.922	63	9592	31.126	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	62.260%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	23469	16.829	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	67.320%	
66) 1,2-Dichlorobenzene-d4	13.849	152	40206	21.867	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	87.480%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100824\
Data File : VW030465.D
Acq On : 08 Oct 2024 23:23
Operator : SY/MD
Sample : VW1008SBL02
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK567

Quant Time: Oct 09 04:19:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Oct 09 01:13:46 2024
Response via : Initial Calibration

