

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100824\
 Data File : VW030452.D
 Acq On : 08 Oct 2024 17:23
 Operator : SY/MD
 Sample : P4329-03
 Misc : 5.00g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4EG4

Quant Time: Oct 09 01:20:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Oct 09 01:13:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	218532	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	168469	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	51298	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	29161	17.077	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	68.320%
7) Chloroethane-d5	2.905	69	22156	17.832	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.320%
11) 1,1-Dichloroethene-d2	4.027	65	23540	18.885	ug/L	0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	75.520%
21) 2-Butanone-d5	7.075	46	14399	20.967	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	41.940%
24) Chloroform-d	7.648	84	106676	18.170	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	72.680%
26) 1,2-Dichloroethane-d4	8.307	65	52431	17.110	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	68.440%#
32) Benzene-d6	8.276	84	212057	22.458	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.840%
36) 1,2-Dichloropropane-d6	9.276	67	54128	19.060	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	76.240%
41) Toluene-d8	10.325	98	205333	23.715	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.880%
43) trans-1,3-Dichloroprop...	10.575	79	22276	18.858	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	75.440%
47) 2-Hexanone-d5	10.922	63	14348	30.596	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	61.200%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	31848	15.008	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	60.040%
66) 1,2-Dichlorobenzene-d4	13.854	152	45265	26.577	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	106.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100824\
Data File : VW030452.D
Acq On : 08 Oct 2024 17:23
Operator : SY/MD
Sample : P4329-03
Misc : 5.00g/10mL/MSVOA_W/SOIL/B
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4EG4

Quant Time: Oct 09 01:20:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Oct 09 01:13:46 2024
Response via : Initial Calibration

