

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100824\
 Data File : VW030456.D
 Acq On : 08 Oct 2024 18:59
 Operator : SY/MD
 Sample : P4329-15
 Misc : 5.59g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4FB4

Quant Time: Oct 09 01:21:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Oct 09 01:13:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	235264	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	199774	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	93221	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	28767	15.648	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	62.600%
7) Chloroethane-d5	2.899	69	22220	16.612	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	66.440%
11) 1,1-Dichloroethene-d2	4.021	65	22713	16.925	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.720%
21) 2-Butanone-d5	7.075	46	10575	14.304	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	28.600%
24) Chloroform-d	7.655	84	107332	16.982	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	67.920%
26) 1,2-Dichloroethane-d4	8.307	65	52099	15.792	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	63.160%#
32) Benzene-d6	8.276	84	213226	19.044	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	76.160%
36) 1,2-Dichloropropane-d6	9.276	67	54448	16.168	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	64.680%#
41) Toluene-d8	10.325	98	206852	20.147	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	80.600%
43) trans-1,3-Dichloroprop...	10.575	79	21769	15.541	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	62.160%
47) 2-Hexanone-d5	10.922	63	10267	18.463	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	36.920%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	35006	13.911	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	55.640%
66) 1,2-Dichlorobenzene-d4	13.849	152	60352	19.500	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	78.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100824\
Data File : VW030456.D
Acq On : 08 Oct 2024 18:59
Operator : SY/MD
Sample : P4329-15
Misc : 5.59g/10mL/MSVOA_W/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4FB4

Quant Time: Oct 09 01:21:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Oct 09 01:13:46 2024
Response via : Initial Calibration

