

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100924\
 Data File : VW030482.D
 Acq On : 09 Oct 2024 18:47
 Operator : SY/MD
 Sample : P4346-08
 Misc : 3.21g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EOAW9MS

Quant Time: Oct 10 01:30:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 10 01:29:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	196459	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	161048	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	54017	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	57779	18.711	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.840%
7) Chloroethane-d5	2.905	69	66039	28.252	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	113.000%
11) 1,1-Dichloroethene-d2	4.021	65	26886	20.933	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	83.720%
21) 2-Butanone-d5	7.081	46	49034	88.233	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	176.460%#
24) Chloroform-d	7.648	84	170928	30.423	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	121.680%
26) 1,2-Dichloroethane-d4	8.307	65	128338	43.928	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	175.720%#
32) Benzene-d6	8.276	84	291439	31.179	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	124.720%
36) 1,2-Dichloropropane-d6	9.276	67	111510	40.645	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	162.600%#
41) Toluene-d8	10.325	98	195021	22.855	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.440%
43) trans-1,3-Dichloroprop...	10.575	79	34112	29.694	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	118.760%
47) 2-Hexanone-d5	10.922	63	38752	109.667	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	219.340%#
56) 1,1,2,2-Tetrachloroeth...	12.690	84	97294	48.289	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	193.160%#
66) 1,2-Dichlorobenzene-d4	13.849	152	58210	31.849	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	127.400%#
Target Compounds						Qvalue
13) Acetone	4.131	43	6709	8.246	ug/L	76
15) Methyl Acetate	4.673	43	5933	5.903	ug/L #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100924\
Data File : VW030482.D
Acq On : 09 Oct 2024 18:47
Operator : SY/MD
Sample : P4346-08
Misc : 3.21g/10mL/MSVOA_W/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EOAW9MS

Quant Time: Oct 10 01:30:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 10 01:29:22 2024
Response via : Initial Calibration

