

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100924\
 Data File : VW030484.D
 Acq On : 09 Oct 2024 19:35
 Operator : SY/MD
 Sample : P4346-10
 Misc : 2.89g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EOAX0

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 10/10/2024
 Supervised By :Mahesh Dadoda 10/10/2024

Quant Time: Oct 10 01:31:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 10 01:29:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	5299	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.629	117	5084	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	1705	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	2004	24.060	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.240%
7) Chloroethane-d5	2.899	69	2186	34.672	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	138.680%
11) 1,1-Dichloroethene-d2	4.039	65	962	27.769	ug/L	0.02
Spiked Amount	25.000	Range	45 - 110	Recovery	=	111.080%#
21) 2-Butanone-d5	7.100	46	2069m	138.030	ug/L	0.02
Spiked Amount	50.000	Range	20 - 135	Recovery	=	276.060%#
24) Chloroform-d	7.654	84	6129	40.445	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	161.760%#
26) 1,2-Dichloroethane-d4	8.313	65	5801	73.616	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	294.480%#
32) Benzene-d6	8.282	84	9656	32.724	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	130.880%
36) 1,2-Dichloropropane-d6	9.276	67	4240	48.957	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	195.840%#
41) Toluene-d8	10.325	98	5897	21.892	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.560%
43) trans-1,3-Dichloroprop...	10.581	79	1125	31.021	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	124.080%
47) 2-Hexanone-d5	10.934	63	778m	69.745	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	139.480%#
56) 1,1,2,2-Tetrachloroeth...	12.690	84	4644	73.014	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	292.040%#
66) 1,2-Dichlorobenzene-d4	13.855	152	2049	35.517	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	142.080%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100924\
Data File : VW030484.D
Acq On : 09 Oct 2024 19:35
Operator : SY/MD
Sample : P4346-10
Misc : 2.89g/10mL/MSVOA_W/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EOAX0

Quant Time: Oct 10 01:31:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 10 01:29:22 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 10/10/2024
Supervised By :Mahesh Dadoda 10/10/2024

