

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100924\
 Data File : VW030481.D
 Acq On : 09 Oct 2024 18:23
 Operator : SY/MD
 Sample : P4346-07
 Misc : 3.42g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EOAW9

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 10/10/2024
 Supervised By :Mahesh Dadoda 10/10/2024

Quant Time: Oct 10 01:30:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 10 01:29:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	308643	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	200039	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	54179	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	142175	29.306	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 117.240%			
7) Chloroethane-d5	2.899	69	126588	34.471	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 137.880%			
11) 1,1-Dichloroethene-d2	4.021	65	55091	27.302	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery = 109.200%			
21) 2-Butanone-d5	7.075	46	55969	64.106	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 128.220%			
24) Chloroform-d	7.648	84	247284	28.016	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery = 112.080%			
26) 1,2-Dichloroethane-d4	8.307	65	165334m	36.022	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery = 144.080%#			
32) Benzene-d6	8.276	84	437563	37.688	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery = 150.760%#			
36) 1,2-Dichloropropane-d6	9.270	67	152190	44.660	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery = 178.640%#			
41) Toluene-d8	10.325	98	272326	25.694	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery = 102.760%			
43) trans-1,3-Dichloroprop...	10.575	79	41097	28.801	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery = 115.200%			
47) 2-Hexanone-d5	10.922	63	45454	103.561	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 207.120%#			
56) 1,1,2,2-Tetrachloroeth...	12.690	84	109242	43.651	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery = 174.600%#			
66) 1,2-Dichlorobenzene-d4	13.849	152	53896	29.400	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery = 117.600%			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100924\
Data File : VW030481.D
Acq On : 09 Oct 2024 18:23
Operator : SY/MD
Sample : P4346-07
Misc : 3.42g/10mL/MSVOA_W/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EOAW9

Quant Time: Oct 10 01:30:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 10 01:29:22 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 10/10/2024
Supervised By :Mahesh Dadoda 10/10/2024

