



Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100924\
Data File : VW030470.D
Acq On : 09 Oct 2024 12:29
Operator : SY/MD
Sample : BFB572
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB572

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Title : SFAM01.0
Last Update : Fri Oct 11 02:35:02 2024



AutoFind: Scans 1759, 1760, 1761; Background Corrected with Scan 1751

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.8	25397	PASS
75	95	30	80	56.5	62872	PASS
95	95	100	100	100.0	111309	PASS
96	95	5	9	6.8	7605	PASS
173	174	0.00	2	1.3	1012	PASS
174	95	50	120	71.0	79029	PASS
175	174	5	9	7.9	6216	PASS
176	174	95	101	96.5	76269	PASS
177	176	5	9	7.2	5477	PASS

m/z	Abundance
35.80	55.0
37.60	87.0
38.70	60.0
40.00	257.0
43.90	177.0
44.60	74.0
48.90	53.0
59.00	60.0
67.10	59.0
69.20	62.0
73.30	60.0

Instrument :
MSVOA_W
ClientSampleId :
BFB572

m/z	Abundance
36.00	1263.0
37.10	6732.0
38.10	6237.0
39.10	2617.0
40.20	272.0
41.10	184.0
43.10	172.0
44.00	1303.0
45.10	1273.0
46.20	199.0
47.10	1370.0
48.00	903.0
49.10	5410.0
50.10	24584.0
51.10	7253.0
52.00	263.0
53.00	276.0
55.00	467.0
56.10	1977.0
57.10	3707.0
58.30	247.0
58.90	193.0
60.10	1301.0
61.00	6551.0
62.00	6321.0
63.10	5509.0
64.00	859.0
65.10	872.0
66.00	204.0
67.00	338.0
68.10	12541.0
69.10	12779.0
70.10	1065.0
71.20	69.0
72.00	602.0
73.10	6139.0
74.10	21808.0
75.10	59168.0
76.10	5323.0
77.10	843.0
79.00	4915.0
80.00	1288.0
81.00	5231.0
81.90	1072.0
83.10	178.0
86.00	278.0
87.00	4442.0
88.00	4862.0
91.00	618.0
92.00	3336.0
93.00	5010.0
94.10	13513.0
95.00	102416.0
96.10	6756.0
97.00	224.0
102.20	56.0
103.00	291.0
104.00	673.0
104.80	261.0
105.90	586.0
106.90	90.0
108.00	60.0
109.60	65.0
110.10	150.0
111.10	247.0
111.80	104.0
112.90	126.0
115.00	173.0
116.00	596.0
117.10	885.0
118.00	523.0
119.00	1051.0
124.00	88.0
125.10	290.0
126.20	88.0
126.60	64.0
127.90	338.0
129.00	198.0
130.00	459.0
130.90	188.0
133.00	782.0
135.10	251.0
136.00	53.0
137.00	185.0
140.10	85.0
140.90	1006.0
142.00	161.0
143.00	1668.0
144.10	182.0
144.90	447.0
146.00	131.0
146.80	122.0

Instrument :
MSVOA_W
ClientSampleId :
BFB572

148.10	159.0
148.90	161.0
150.00	122.0
154.90	230.0
156.90	167.0
158.00	83.0
158.90	111.0
160.90	189.0
163.00	101.0
165.00	125.0
169.00	86.0
171.00	205.0
172.00	1411.0
173.10	1694.0
174.00	70656.0
175.00	5129.0
175.90	51976.0
177.00	4436.0
177.90	265.0
179.00	290.0
179.90	62.0
189.10	61.0
191.10	603.0
192.20	244.0
193.10	789.0
193.90	121.0
194.40	143.0
194.90	101.0
203.20	55.0
205.10	172.0
205.60	52.0
207.00	527.0
234.90	98.0
249.10	449.0
250.00	173.0
251.10	278.0
252.20	80.0
265.10	778.0
266.20	227.0
266.90	222.0
267.70	56.0
268.20	50.0
279.80	182.0
281.10	7297.0
282.10	2614.0
283.00	1019.0
283.90	212.0
285.00	109.0

Instrument :
MSVOA_W
ClientSampleId :
BFB572

m/z	Abundance
36.10	1696.0
37.10	7629.0
38.10	7141.0
39.00	2969.0
40.10	348.0
40.90	141.0
41.50	60.0
42.30	50.0
43.10	186.0
44.10	1176.0
45.00	1189.0
46.10	222.0
47.00	1354.0
48.10	1018.0
49.10	5957.0
50.10	29304.0
51.10	8387.0
52.30	272.0
53.00	210.0
54.20	97.0
55.10	465.0
56.00	2540.0
57.10	4493.0
58.00	219.0
58.70	166.0
60.00	1196.0
61.10	7630.0
62.10	7093.0
63.10	6464.0
64.00	850.0
65.10	1167.0
66.10	197.0
67.00	499.0
68.00	14507.0
69.10	14293.0
70.10	1111.0
71.20	108.0
72.00	910.0
73.10	7243.0
74.10	24968.0
75.10	69744.0
76.10	6036.0
77.00	801.0
78.00	635.0
79.00	6320.0
80.00	1694.0
81.00	6016.0
82.00	972.0
83.10	245.0
85.20	109.0
85.80	156.0
87.00	5178.0
88.00	5470.0
89.90	75.0
91.00	689.0
92.10	4580.0
93.10	5993.0
94.00	16808.0
95.10	125168.0
96.10	8926.0
97.10	260.0
102.20	63.0
103.10	198.0
104.00	769.0
105.00	334.0
105.90	709.0
107.00	230.0
108.00	66.0
110.00	117.0
111.00	145.0
111.90	181.0
112.80	185.0
114.90	177.0
116.00	567.0
117.00	1092.0
117.90	559.0
119.00	880.0
123.10	59.0
123.90	105.0
125.00	290.0
126.10	100.0
126.90	84.0
128.10	543.0
128.90	190.0
129.90	480.0
131.00	182.0
133.00	385.0
134.00	181.0
135.00	264.0
135.90	74.0
136.90	353.0
139.00	63.0

Instrument :
MSVOA_W
ClientSampleId :
BFB572

140.20	153.0
140.90	1498.0
141.90	265.0
143.00	1779.0
145.00	485.0
146.00	295.0
146.90	129.0
147.90	385.0
148.90	124.0
150.00	226.0
152.00	108.0
153.00	74.0
153.80	109.0
155.00	328.0
156.80	215.0
158.90	210.0
160.80	143.0
163.10	53.0
165.10	58.0
170.20	68.0
171.00	334.0
172.00	1545.0
174.00	83056.0
175.00	7361.0
176.00	88032.0
177.00	6551.0
178.10	195.0
178.90	127.0
189.10	77.0
191.10	372.0
192.10	124.0
192.90	485.0
194.00	117.0
194.90	101.0
203.20	62.0
205.00	64.0
206.90	202.0
208.30	138.0
249.10	312.0
250.00	107.0
250.90	118.0
264.90	402.0
266.10	59.0
267.10	71.0
281.10	3981.0
282.10	926.0
283.20	671.0
284.20	154.0

Instrument :
MSVOA_W
ClientSampleId :
BFB572

m/z	Abundance
36.10	1056.0
37.10	5863.0
38.10	5605.0
39.10	2152.0
39.90	241.0
40.20	230.0
41.40	144.0
42.20	72.0
42.90	181.0
44.00	1054.0
45.00	1254.0
46.30	228.0
47.00	1257.0
48.10	711.0
49.10	5002.0
50.10	22304.0
51.10	6632.0
52.00	316.0
52.80	157.0
53.20	107.0
53.90	166.0
55.10	331.0
56.00	1966.0
57.10	3418.0
58.00	141.0
58.80	145.0
60.00	1028.0
61.00	5684.0
62.10	6030.0
63.10	5089.0
64.00	753.0
65.10	970.0
66.00	118.0
67.10	386.0
68.10	11924.0
69.10	11768.0
70.10	936.0
72.10	762.0
73.10	5380.0
74.10	20816.0
75.10	59704.0
76.10	4842.0
77.10	742.0
77.90	485.0
79.00	5000.0
80.00	1282.0
81.00	4830.0
82.00	1193.0
83.00	130.0
83.40	71.0
85.90	275.0
87.00	4765.0
88.00	3880.0
91.10	613.0
92.00	3692.0
93.10	5262.0
94.10	13481.0
95.10	106344.0
96.10	7132.0
97.00	175.0
102.90	122.0
104.00	671.0
105.00	338.0
105.90	641.0
107.00	127.0
108.10	51.0
110.20	229.0
111.10	148.0
111.80	171.0
113.10	177.0
114.90	186.0
116.10	548.0
116.90	855.0
117.90	381.0
118.90	894.0
123.90	143.0
125.00	130.0
126.00	113.0
126.90	57.0
128.00	517.0
128.90	213.0
129.90	595.0
130.90	227.0
133.20	158.0
134.20	60.0
134.90	232.0
135.90	62.0
136.80	252.0
140.10	116.0
140.90	1321.0
141.90	122.0
142.90	1412.0

Instrument :
MSVOA_W
ClientSampleId :
BFB572

144.00	102.0
145.10	182.0
146.00	268.0
147.00	77.0
148.00	319.0
149.10	81.0
150.00	155.0
152.00	92.0
153.10	92.0
153.70	88.0
154.80	282.0
156.20	53.0
157.10	301.0
158.90	160.0
160.80	110.0
163.30	59.0
169.90	56.0
172.00	1091.0
172.90	1342.0
174.00	83376.0
175.00	6159.0
176.00	88800.0
177.00	5445.0
178.10	223.0
178.90	77.0
191.10	189.0
192.00	57.0
193.00	206.0
194.10	57.0
207.00	171.0
208.20	63.0
249.00	156.0
250.90	65.0
251.20	62.0
264.80	143.0
265.30	285.0
265.80	54.0
267.00	60.0
279.60	72.0
281.20	2049.0
282.00	718.0
283.10	301.0

Instrument :
MSVOA_W
ClientSampleId :
BFB572