

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101022\
 Data File : VW024811.D
 Acq On : 10 Oct 2022 16:56
 Operator : SY/VA
 Sample : N5049-15
 Misc : 6.41g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0BS1

Quant Time: Oct 11 01:59:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 07 03:04:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	525322	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	448242	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	182515	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	73160	13.311	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	53.240%
7) Chloroethane-d5	2.885	69	52623	16.476	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.920%
11) 1,1-Dichloroethene-d2	4.019	63	115169	9.053	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	36.200%#
21) 2-Butanone-d5	7.080	46	120207	60.832	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	121.660%
24) Chloroform-d	7.653	84	316864	22.563	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.240%
26) 1,2-Dichloroethane-d4	8.305	65	173453	23.809	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	95.240%
32) Benzene-d6	8.275	84	498006	17.382	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	69.520%
36) 1,2-Dichloropropane-d6	9.274	67	198180	22.640	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	90.560%
41) Toluene-d8	10.323	98	365081	14.257	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	57.040%
43) trans-1,3-Dichloroprop...	10.579	79	54184	13.900	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	55.600%
47) 2-Hexanone-d5	10.927	63	81438	54.375	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	108.740%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	194231	31.141	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	124.560%#
66) 1,2-Dichlorobenzene-d4	13.847	152	123013	19.153	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	76.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101022\
Data File : VW024811.D
Acq On : 10 Oct 2022 16:56
Operator : SY/VA
Sample : N5049-15
Misc : 6.41g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0BS1

Quant Time: Oct 11 01:59:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 07 03:04:02 2022
Response via : Initial Calibration

