

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101022\
 Data File : VW024832.D
 Acq On : 11 Oct 2022 02:35
 Operator : SY/VA
 Sample : N5064-08
 Misc : 7.09g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0BP4

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/11/2022
 Supervised By :Mahesh Dadoda 10/11/2022

Quant Time: Oct 11 06:37:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 07 03:04:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	533034	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	485248	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	214206	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	105109	18.847	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	75.400%		
7) Chloroethane-d5	2.885	69	63381	19.557	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	78.240%		
11) 1,1-Dichloroethene-d2	4.019	63	166218	12.877	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	51.520%		
21) 2-Butanone-d5	7.080	46	124541	62.114	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	124.220%		
24) Chloroform-d	7.647	84	321474	22.560	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	90.240%		
26) 1,2-Dichloroethane-d4	8.305	65	182271m	24.657	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	98.640%		
32) Benzene-d6	8.268	84	625930	20.181	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	80.720%		
36) 1,2-Dichloropropane-d6	9.274	67	208856	22.040	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	88.160%		
41) Toluene-d8	10.323	98	533603	19.249	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	77.000%		
43) trans-1,3-Dichloroprop...	10.579	79	74343	17.616	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	70.480%		
47) 2-Hexanone-d5	10.920	63	96200	59.333	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	118.660%		
56) 1,1,2,2-Tetrachloroeth...	12.688	84	193453	28.651	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	114.600%		
66) 1,2-Dichlorobenzene-d4	13.847	152	164564	21.832	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	87.320%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101022\
Data File : VW024832.D
Acq On : 11 Oct 2022 02:35
Operator : SY/VA
Sample : N5064-08
Misc : 7.09g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0BP4

Quant Time: Oct 11 06:37:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 07 03:04:02 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 10/11/2022
Supervised By : Mahesh Dadoda 10/11/2022

