

Data File : VW013547.D
Acq On : 11 Oct 2019 19:14
Operator : SY/VA
Sample : K5262-14RE
Misc : 3.69G/10ML/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFBB0RE

Quant Time: Oct 11 22:35:45 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM101019S.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 11 22:30:37 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	0.00	114	0	0.00	ug/L	-8.84
28) Chlorobenzene-d5	0.00	117	0	0.00	ug/L	-11.63
60) 1,4-Dichlorobenzene-d4	0.00	152	0	0.00	ug/L	-13.56

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	22	0.00	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	0.00%#
7) Chloroethane-d5	2.90	69	68	0.00	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	0.00%#
10) 1,1-Dichloroethene-d2	4.03	63	49	0.00	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	0.00%#
20) 2-Butanone-d5	0.00	46	0	0.00	ug/L	
Spiked Amount	50.000	Range	20 - 135	Recovery	=	0.00%#
24) Chloroform-d	7.72	84	53	0.00	ug/L	0.07
Spiked Amount	25.000	Range	40 - 150	Recovery	=	0.00%#
26) 1,2-Dichloroethane-d4	0.00	65	0	0.00	ug/L	
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.00%#
29) Benzene-d6	8.30	84	93	0.00	ug/L	0.02
Spiked Amount	25.000	Range	20 - 135	Recovery	=	0.00%#
33) 1,2-Dichloropropane-d6	0.00	67	0	0.00	ug/L	
Spiked Amount	25.000	Range	70 - 120	Recovery	=	0.00%#
37) Toluene-d8	0.00	98	0	0.00	ug/L	
Spiked Amount	25.000	Range	30 - 130	Recovery	=	0.00%#
38) trans-1,3-Dichloropropene-	0.00	79	0	0.00	ug/L	
Spiked Amount	25.000	Range	30 - 135	Recovery	=	0.00%#
39) 2-Hexanone-d5	0.00	63	0	0.00	ug/L	
Spiked Amount	50.000	Range	20 - 135	Recovery	=	0.00%#
48) 1,1,2,2-Tetrachloroethane-	0.00	84	0	0.00	ug/L	
Spiked Amount	25.000	Range	45 - 120	Recovery	=	0.00%#
61) 1,2-Dichlorobenzene-d4	0.00	152	0	0.00	ug/L	
Spiked Amount	25.000	Range	75 - 120	Recovery	=	0.00%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VW013547.D
Acq On : 11 Oct 2019 19:14
Operator : SY/VA
Sample : K5262-14RE
Misc : 3.69G/10ML/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFBB0RE

Quant Time: Oct 11 22:35:45 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM101019S.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 11 22:30:37 2019
Response via : Initial Calibration

