

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW101119\
 Data File : VW013538.D
 Acq On : 11 Oct 2019 15:20
 Operator : SY/VA
 Sample : K5265-22
 Misc : 6.89G/10ML/MSVOA W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBAT4

Quant Time: Oct 11 23:23:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM101019S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 11 22:30:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	432885	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	365160	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.55	152	154554	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	91226	18.42	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.68%
7) Chloroethane-d5	2.89	69	80611	21.19	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.76%
10) 1,1-Dichloroethene-d2	4.02	63	147835	14.38	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.52%
20) 2-Butanone-d5	7.08	46	73279	49.10	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.20%
24) Chloroform-d	7.65	84	220676	23.47	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.88%
26) 1,2-Dichloroethane-d4	8.30	65	116092	23.83	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	95.32%
29) Benzene-d6	8.27	84	454960	23.26	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.04%
33) 1,2-Dichloropropane-d6	9.27	67	145301	25.04	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.16%
37) Toluene-d8	10.32	98	415638	22.39	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	89.56%
38) trans-1,3-Dichloropropene-	10.58	79	54654	21.14	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.56%
39) 2-Hexanone-d5	10.92	63	57548	49.57	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.14%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	114391	24.59	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	98.36%
61) 1,2-Dichlorobenzene-d4	13.85	152	136188	24.52	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.08%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW101119\
Data File : VW013538.D
Acq On : 11 Oct 2019 15:20
Operator : SY/VA
Sample : K5265-22
Misc : 6.89G/10ML/MSVOA W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBAT4

Quant Time: Oct 11 23:23:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM101019S.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 11 22:30:37 2019
Response via : Initial Calibration

