

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101122\
 Data File : VW024846.D
 Acq On : 11 Oct 2022 13:16
 Operator : SY/VA
 Sample : N5064-14
 Misc : 7.47g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0BH7

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 10/12/2022
 Supervised By : Mahesh Dadoda 10/12/2022

Quant Time: Oct 12 01:20:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Oct 12 01:17:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	485139	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	419978	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	181241	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	110639	21.797	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	87.200%		
7) Chloroethane-d5	2.886	69	67247	22.799	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	91.200%		
11) 1,1-Dichloroethene-d2	4.019	63	179827	15.307	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	61.240%		
21) 2-Butanone-d5	7.080	46	119398	65.427	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	130.860%		
24) Chloroform-d	7.653	84	305507	23.556	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	94.240%		
26) 1,2-Dichloroethane-d4	8.305	65	179821m	26.727	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	106.920%		
32) Benzene-d6	8.275	84	593049	22.093	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	88.360%		
36) 1,2-Dichloropropane-d6	9.275	67	199260	24.295	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	97.200%		
41) Toluene-d8	10.323	98	492374	20.523	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	82.080%		
43) trans-1,3-Dichloroprop...	10.579	79	68345	18.712	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	74.840%		
47) 2-Hexanone-d5	10.921	63	87109	62.075	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	124.160%		
56) 1,1,2,2-Tetrachloroeth...	12.689	84	187829	32.141	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	128.560%#		
66) 1,2-Dichlorobenzene-d4	13.853	152	146440	22.961	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	91.840%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101122\
Data File : VW024846.D
Acq On : 11 Oct 2022 13:16
Operator : SY/VA
Sample : N5064-14
Misc : 7.47g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0BH7

Quant Time: Oct 12 01:20:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Oct 12 01:17:56 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/12/2022
Supervised By :Mahesh Dadoda 10/12/2022

