

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101122\
 Data File : VW024847.D
 Acq On : 11 Oct 2022 13:40
 Operator : SY/VA
 Sample : N5064-15
 Misc : 7.34g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 COBM3

Quant Time: Oct 12 01:20:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Oct 12 01:17:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	572607	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	523635	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	234962	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	118986	19.861	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.440%
7) Chloroethane-d5	2.885	69	71360	20.497	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.000%
11) 1,1-Dichloroethene-d2	4.019	63	200543	14.463	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.840%
21) 2-Butanone-d5	7.080	46	124965	58.018	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	116.040%
24) Chloroform-d	7.647	84	349604	22.838	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.360%
26) 1,2-Dichloroethane-d4	8.305	65	198182	24.957	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	99.840%
32) Benzene-d6	8.275	84	687436	20.540	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	82.160%
36) 1,2-Dichloropropane-d6	9.274	67	228995	22.393	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	89.560%
41) Toluene-d8	10.323	98	579475	19.372	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	77.480%
43) trans-1,3-Dichloroprop...	10.579	79	79875	17.540	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	70.160%
47) 2-Hexanone-d5	10.926	63	94712	54.133	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	108.260%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	209932	28.812	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	115.240%
66) 1,2-Dichlorobenzene-d4	13.853	152	181000	21.892	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	87.560%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101122\
Data File : VW024847.D
Acq On : 11 Oct 2022 13:40
Operator : SY/VA
Sample : N5064-15
Misc : 7.34g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0BM3

Quant Time: Oct 12 01:20:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Oct 12 01:17:56 2022
Response via : Initial Calibration

