

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101122\
 Data File : VW024849.D
 Acq On : 11 Oct 2022 14:28
 Operator : SY/VA
 Sample : N5064-17
 Misc : 7.20g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 COBN6

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 10/12/2022
 Supervised By : Mahesh Dadoda 10/12/2022

Quant Time: Oct 12 01:21:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Oct 12 01:17:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	1461	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.640	117	1140m	25.000	ug/L	0.01
58) 1,4-Dichlorobenzene-d4	13.560	152	289	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	1023	66.924	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	267.680%#
7) Chloroethane-d5	2.904	69	924	104.022	ug/L	0.02
Spiked Amount	25.000	Range	30 - 150	Recovery	=	416.080%#
11) 1,1-Dichloroethene-d2	4.038	63	1153m	32.589	ug/L	0.02
Spiked Amount	25.000	Range	45 - 110	Recovery	=	130.360%#
21) 2-Butanone-d5	0.000	46	0d	0.000	ug/L	
Spiked Amount	50.000	Range	20 - 135	Recovery	=	0.000%#
24) Chloroform-d	7.647	84	1595	40.837	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	163.360%#
26) 1,2-Dichloroethane-d4	8.311	65	784	38.694	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	154.760%#
32) Benzene-d6	8.281	84	2352	32.279	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	129.120%
36) 1,2-Dichloropropane-d6	9.287	67	814m	36.563	ug/L	0.01
Spiked Amount	25.000	Range	70 - 120	Recovery	=	146.240%#
41) Toluene-d8	10.335	98	1618	24.845	ug/L	0.01
Spiked Amount	25.000	Range	30 - 130	Recovery	=	99.360%
43) trans-1,3-Dichloroprop...	0.000	79	0	0.000	ug/L	
Spiked Amount	25.000	Range	30 - 135	Recovery	=	0.000%#
47) 2-Hexanone-d5	0.000	63	0d	0.000	ug/L	
Spiked Amount	50.000	Range	20 - 135	Recovery	=	0.000%#
56) 1,1,2,2-Tetrachloroeth...	12.695	84	616	38.833	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	155.320%#
66) 1,2-Dichlorobenzene-d4	13.853	152	177	17.405	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	69.600%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101122\
Data File : VW024849.D
Acq On : 11 Oct 2022 14:28
Operator : SY/VA
Sample : N5064-17
Misc : 7.20g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0BN6

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 10/12/2022
Supervised By : Mahesh Dadoda 10/12/2022

Quant Time: Oct 12 01:21:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Oct 12 01:17:56 2022
Response via : Initial Calibration

