

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101124\
 Data File : VW030556.D
 Acq On : 11 Oct 2024 18:02
 Operator : SY/MD
 Sample : P4350-05
 Misc : 5.12g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Oct 14 01:44:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Oct 14 01:34:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	606450	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	548382	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.550	152	243954	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	200997	21.085	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.360%
7) Chloroethane-d5	2.899	69	161534	22.387	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.560%
11) 1,1-Dichloroethene-d2	4.027	65	88726	22.378	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	89.520%
21) 2-Butanone-d5	7.075	46	85756	49.989	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.980%
24) Chloroform-d	7.734	84	357	0.021	ug/L	0.09
Spiked Amount	25.000	Range	40 - 150	Recovery	=	0.080%#
26) 1,2-Dichloroethane-d4	8.307	65	222461	24.667	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	98.680%
32) Benzene-d6	8.270	84	752737	23.650	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	94.600%
36) 1,2-Dichloropropane-d6	9.270	67	228483	24.458	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	97.840%
41) Toluene-d8	10.325	98	691501	23.799	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	95.200%
43) trans-1,3-Dichloroprop...	10.581	79	96877	24.766	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	99.080%
47) 2-Hexanone-d5	10.922	63	70587	58.665	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	117.340%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	176178	25.680	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	102.720%
66) 1,2-Dichlorobenzene-d4	13.848	152	223528	27.080	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	108.320%
Target Compounds						Qvalue
13) Acetone	4.118	43	22714	9.044	ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101124\
Data File : VW030556.D
Acq On : 11 Oct 2024 18:02
Operator : SY/MD
Sample : P4350-05
Misc : 5.12g/10mL/MSVOA_W/SOIL/A
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :


Quant Time: Oct 14 01:44:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Oct 14 01:34:28 2024
Response via : Initial Calibration





#13
 Acetone
 Concen: 9.044 ug/L
 RT: 4.118 min Scan# 3
 Delta R.T. -0.000 min
 Lab File: VW030556.D
 Acq: 11 Oct 2024 18:02

Instrument :
 MSVOA_W
 ClientSampleId :



Tgt Ion: 43 Resp: 22714
 Ion Ratio Lower Upper
 43 100
 58 30.4 0.0 57.6

