

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101124\
 Data File : VW030557.D
 Acq On : 11 Oct 2024 18:26
 Operator : SY/MD
 Sample : P4350-06
 Misc : 5.68g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EOAK3

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 10/15/2024
 Supervised By :Semsettin Yesilyurt 10/15/2024

Quant Time: Oct 14 01:44:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Oct 14 01:34:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.849	114	3718	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.641	117	2010	25.000	ug/L	0.01
58) 1,4-Dichlorobenzene-d4	13.568	152	476m	25.000	ug/L	0.01
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	1714	29.329	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	117.320%
7) Chloroethane-d5	2.905	69	1270	28.709	ug/L	0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	114.840%
11) 1,1-Dichloroethene-d2	4.033	65	807	33.200	ug/L	0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	132.800%#
21) 2-Butanone-d5	0.000	46	0d	0.000	ug/L	
Spiked Amount	50.000	Range	20 - 135	Recovery	=	0.000%#
24) Chloroform-d	7.660	84	2514	23.644	ug/L	0.01
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.560%
26) 1,2-Dichloroethane-d4	8.307	65	1289m	23.313	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	93.240%
32) Benzene-d6	8.276	84	5377	46.091	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	184.360%#
36) 1,2-Dichloropropane-d6	9.276	67	1764	51.517	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	206.080%#
41) Toluene-d8	10.325	98	4020	37.747	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	151.000%#
43) trans-1,3-Dichloroprop...	0.000	79	0d	0.000	ug/L	
Spiked Amount	25.000	Range	30 - 135	Recovery	=	0.000%#
47) 2-Hexanone-d5	0.000	63	0d	0.000	ug/L	
Spiked Amount	50.000	Range	20 - 135	Recovery	=	0.000%#
56) 1,1,2,2-Tetrachloroeth...	12.690	84	556	22.111	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	88.440%
66) 1,2-Dichlorobenzene-d4	13.848	152	345	21.421	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	85.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101124\
Data File : VW030557.D
Acq On : 11 Oct 2024 18:26
Operator : SY/MD
Sample : P4350-06
Misc : 5.68g/10mL/MSVOA_W/SOIL/A
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EOAK3

Quant Time: Oct 14 01:44:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Oct 14 01:34:28 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 10/15/2024
Supervised By :Semsettin Yesilyurt 10/15/2024

