

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101124\
 Data File : VW030571.D
 Acq On : 12 Oct 2024 01:13
 Operator : SY/MD
 Sample : P4361-11
 Misc : 3.34g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EOAZ2

Quant Time: Oct 14 04:50:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Oct 14 04:33:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	443330	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	359242	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.555	152	146334	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	165839	23.798	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	95.200%
7) Chloroethane-d5	2.899	69	137205	26.011	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	104.040%
11) 1,1-Dichloroethene-d2	4.021	65	63755	21.997	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	88.000%
21) 2-Butanone-d5	7.087	46	57780	46.074	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.140%
24) Chloroform-d	7.648	84	281852	22.231	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	88.920%
26) 1,2-Dichloroethane-d4	8.306	65	160114	24.286	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.160%
32) Benzene-d6	8.276	84	529615	25.401	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	101.600%
36) 1,2-Dichloropropane-d6	9.270	67	162400	26.537	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	106.160%
41) Toluene-d8	10.324	98	435270	22.868	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.480%
43) trans-1,3-Dichloroprop...	10.574	79	59284	23.135	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	92.520%
47) 2-Hexanone-d5	10.922	63	46138	58.534	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	117.060%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	114098	25.387	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	101.560%
66) 1,2-Dichlorobenzene-d4	13.848	152	120323	24.301	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.200%
Target Compounds						
13) Acetone	4.142	43	9129	4.972	ug/L	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101124\
Data File : VW030571.D
Acq On : 12 Oct 2024 01:13
Operator : SY/MD
Sample : P4361-11
Misc : 3.34g/10mL/MSVOA_W/SOIL/A
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EOAZ2

Quant Time: Oct 14 04:50:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Oct 14 04:33:53 2024
Response via : Initial Calibration

