

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101124\
 Data File : VW030551.D
 Acq On : 11 Oct 2024 16:02
 Operator : SY/MD
 Sample : VIBLK545
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK545

Quant Time: Oct 14 01:42:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Oct 14 01:34:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	269771	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	256379	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	112500	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	117454	27.699	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	110.800%
7) Chloroethane-d5	2.899	69	94926	29.574	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	118.280%
11) 1,1-Dichloroethene-d2	4.027	65	49899	28.292	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	113.160%#
21) 2-Butanone-d5	7.075	46	46526	60.969	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	121.940%
24) Chloroform-d	7.648	84	221004	28.646	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	114.600%
26) 1,2-Dichloroethane-d4	8.301	65	120973	30.155	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	120.600%
32) Benzene-d6	8.276	84	421307	28.313	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	113.240%
36) 1,2-Dichloropropane-d6	9.276	67	127946	29.295	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	117.200%
41) Toluene-d8	10.325	98	395189	29.092	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	116.360%
43) trans-1,3-Dichloroprop...	10.581	79	50428	27.574	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	110.280%
47) 2-Hexanone-d5	10.922	63	34358	61.078	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	122.160%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	94854	29.573	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	118.280%
66) 1,2-Dichlorobenzene-d4	13.855	152	123062	32.329	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	129.320%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101124\
Data File : VW030551.D
Acq On : 11 Oct 2024 16:02
Operator : SY/MD
Sample : VIBLK545
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK545

Quant Time: Oct 14 01:42:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM100924SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Oct 14 01:34:28 2024
Response via : Initial Calibration

