

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101322\
 Data File : VW024901.D
 Acq On : 13 Oct 2022 14:10
 Operator : SY/VA
 Sample : N5070-03
 Misc : 4.97g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0BK7

Quant Time: Oct 14 02:12:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 14 02:07:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	437837	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	339598	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	103057	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	94432	18.281	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.120%
7) Chloroethane-d5	2.885	69	58617	19.382	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.520%
11) 1,1-Dichloroethene-d2	4.019	63	155300	13.828	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	55.320%
21) 2-Butanone-d5	7.080	46	85866	52.464	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	104.920%
24) Chloroform-d	7.647	84	284126	23.007	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.040%
26) 1,2-Dichloroethane-d4	8.305	65	166356	25.739	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	102.960%
32) Benzene-d6	8.275	84	536208	25.180	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	100.720%
36) 1,2-Dichloropropane-d6	9.275	67	182277	27.907	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	111.640%
41) Toluene-d8	10.323	98	410520	21.478	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.920%
43) trans-1,3-Dichloroprop...	10.579	79	58665	23.682	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	94.720%
47) 2-Hexanone-d5	10.927	63	68014	62.109	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	124.220%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	153466	30.128	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	120.520%#
66) 1,2-Dichlorobenzene-d4	13.847	152	80883	21.476	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	85.920%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101322\
Data File : VW024901.D
Acq On : 13 Oct 2022 14:10
Operator : SY/VA
Sample : N5070-03
Misc : 4.97g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0BK7

Quant Time: Oct 14 02:12:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101222SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 14 02:07:33 2022
Response via : Initial Calibration

