

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101422\
 Data File : VW024923.D
 Acq On : 14 Oct 2022 11:51
 Operator : SY/VA
 Sample : N5049-17
 Misc : 5.01g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Oct 15 00:16:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Oct 15 00:14:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	456103	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	425255	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	185242	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	110124	20.466	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.880%
7) Chloroethane-d5	2.885	69	63459	20.142	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.560%
11) 1,1-Dichloroethene-d2	4.013	63	172785	14.769	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.080%
21) 2-Butanone-d5	7.080	46	81747	47.947	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.900%
24) Chloroform-d	7.647	84	277877	21.600	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.400%
26) 1,2-Dichloroethane-d4	8.305	65	152481	22.647	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	90.600%
32) Benzene-d6	8.275	84	571808	21.443	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.760%
36) 1,2-Dichloropropane-d6	9.274	67	179667	21.967	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.880%
41) Toluene-d8	10.323	98	511299	21.362	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.440%
43) trans-1,3-Dichloroprop...	10.579	79	69039	22.256	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.040%
47) 2-Hexanone-d5	10.921	63	64686	47.172	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.340%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	142153	22.286	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.160%
66) 1,2-Dichlorobenzene-d4	13.853	152	161044	23.790	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101422\
Data File : VW024923.D
Acq On : 14 Oct 2022 11:51
Operator : SY/VA
Sample : N5049-17
Misc : 5.01g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Oct 15 00:16:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101222SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Oct 15 00:14:44 2022
Response via : Initial Calibration

