

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101422\
 Data File : VW024929.D
 Acq On : 14 Oct 2022 14:17
 Operator : SY/VA
 Sample : N5144-08
 Misc : 3.90g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0BR2

Quant Time: Oct 15 00:17:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Oct 15 00:14:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	490291	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	450887	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	204344	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	102839	17.779	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.120%
7) Chloroethane-d5	2.885	69	66846	19.738	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.960%
11) 1,1-Dichloroethene-d2	4.019	63	173822	13.821	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	55.280%
21) 2-Butanone-d5	7.080	46	119541	65.225	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	130.440%
24) Chloroform-d	7.653	84	317217	22.939	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.760%
26) 1,2-Dichloroethane-d4	8.311	65	174722	24.141	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	96.560%
32) Benzene-d6	8.275	84	602497	21.309	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	85.240%
36) 1,2-Dichloropropane-d6	9.274	67	201010	23.179	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.720%
41) Toluene-d8	10.323	98	510710	20.125	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	80.480%
43) trans-1,3-Dichloroprop...	10.579	79	73382	22.312	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.240%
47) 2-Hexanone-d5	10.920	63	88215	60.673	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	121.340%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	188094	27.812	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	111.240%
66) 1,2-Dichlorobenzene-d4	13.853	152	172844	23.146	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	92.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW101422\
Data File : VW024929.D
Acq On : 14 Oct 2022 14:17
Operator : SY/VA
Sample : N5144-08
Misc : 3.90g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0BR2

Quant Time: Oct 15 00:17:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM101222SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Oct 15 00:14:44 2022
Response via : Initial Calibration

